

GB/US power one 675charger

Scope of delivery

- 1 x **power one 675charger**
- 1 x connection lead for charger / power source connection
- 1 x mains adaptor with USB port for connection lead
- 1 x set of plug attachments for mains adaptor with 4 different international connections

Operating instructions

Dear Customer,

Congratulations on buying a power one 675charger. You have acquired a modern, safe and reliable charging system for power one 675 ACCU plus rechargeable hearing aid batteries. Please read through the following operating instructions carefully, to ensure trouble-free use of the device.

1. Description

The **power one 675charger** is used for charging one or two **power one p675 ACCU plus** rechargeable hearing aid batteries. The device recharges the **power one p675 ACCU plus** within 5 hours (only on previously fully-discharged rechargeable batteries / partially-charged rechargeable batteries are fully discharged first for safety and for improved durability / refreshing protection).

Please use only power one ACCU plus brand rechargeable batteries for the power one 675charger.

2. Use

Please open the package and check that all the components listed in the delivery package are enclosed.

For charging, please take out the connection lead (supplied) and pull it out to the required length (pull on both ends at the same time!). Then connect the coaxial DC power plug to the corresponding connection on the **power one 675charger**, located on the flat reverse side of the charger. Now connect the USB connection of the connection lead to a suitable power source. The USB connection can be connected to a socket outlet using the mains adaptor (supplied). The plug attachments (four different plug attachments in total) are interchangeable and can be selected according to the country/region.

You can also use the charger by connecting it directly to a USB port of a powered-up computer or laptop, via the USB plug. Another possible charging method is to connect the USB plug to the cigar lighter or some other 12V connection in a car, via an adaptor (adaptor not included in delivery package).

The **power one 675charger** is now ready for operation and can be used to charge **power one p675 ACCU plus** nickel-metal hydride hearing aid batteries. To charge the batteries, slide the silver-coloured charger cover towards the rounded side.

Now insert one or two **power one p675 ACCU plus** nickel-metal hydride hearing aid batteries in the battery compartment provided, with the + side (flat side of the button cell) downwards. Ensure correct polarity when inserting the button cells. Very short flashing signals (0.1 sec. on and 0.1 sec. off) are displayed in the event of wrong polarity or a faulty button cell. In this case, no charging takes place. Now close the silver-coloured cover again. The cover must be closed for the batteries to be charged. Charging of the nickel-metal hydride hearing aid batteries is indicated by a green flashing signal.

The **power one 675charger** automatically checks the state of charge of the nickel-metal hydride rechargeable button cells inserted. Button cells discharged in a hearing aid are fully recharged within 5 hours. A green steady signal comes on at the end of charging and you can take out the fully-recharged button cells for use in your hearing aid.

3. Handling instructions for NiMH rechargeable batteries

Please keep the contact surfaces of the rechargeable batteries and the contact surfaces in the device clean at all times. In principle, for the best possible performance and service life, please avoid deep discharge or short circuit of the battery. After a long storage period, full capacity is only restored after several charge/discharge cycles.

Illustrations (see back page)

- 1. power one 675charger**
2. USB connection lead/ coaxial DC power plug
3. Mains adaptor with interchangeable plug attachments
4. Suitable NiMH hearing aid battery – the **power one ACCU plus** (not included in delivery package)